

## New England Biolabs Certificate of Analysis

**Product Name:** Styl-HF®  
**Catalog Number:** R3500S  
**Concentration:** 20,000 U/ml  
**Unit Definition:** One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.  
**Packaging Lot Number:** 10068469  
**Expiration Date:** 02/2022  
**Storage Temperature:** -20°C  
**Storage Conditions:** 50 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA  
**Specification Version:** PS-R3500S/L v1.0

Styl-HF® Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R3500SVIAL	Styl-HF®	10065908	Pass
B7204SVIAL	CutSmart® Buffer	10068805	Pass
B7024SVIAL	Gel Loading Dye, Purple (6X)	10065745	Pass


Assay Name/Specification	Lot # 10068469
<b>Ligation and Recutting (Terminal Integrity)</b> After a 50-fold over-digestion of Lambda DNA with Styl-HF™, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Styl-HF™.	Pass
<b>Exonuclease Activity (Radioactivity Release)</b> A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [ <sup>3</sup> H] E. coli DNA and a minimum of 100 units of Styl-HF™ incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
<b>Endonuclease Activity (Nicking)</b> A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 20 Units of Styl-HF™ incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
<b>Non-Specific DNase Activity (16 Hour)</b> A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 100 Units of Styl-HF™ incubated for 16 hours at 37°C results in a DNA pattern free	Pass

Assay Name/Specification	Lot # 10068469
of detectable nuclease degradation as determined by agarose gel electrophoresis.	

This product has been tested and shown to be in compliance with all specifications.



Anthony Francis  
Production Scientist  
30 Jan 2020



Michael Tonello  
Packaging Quality Control Inspector  
23 Mar 2020