

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name:	DraIII-HFTM
Catalog #:	R3510S/L
Concentration:	20,000 units/ml
Unit Definition:	One unit is defined as the amount of enzyme required to digest 1 μg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 μl.
<i>Lot</i> #:	0021502
Assay Date:	02/2015
Expiration Date:	2/2017
Storage Temp:	-20°C
Storage Conditions:	20 mM Tris-HCl, 300 mM NaCl, 1 mM Dithiothreitol, 0.1 mM EDTA, 200 µg/ml BSA, 50% Glycerol, pH 7.4 @ 25°C
Specification Version:	PS-R3510S/L v1.0
Effective Date:	24 Apr 2013

Assay Name/Specification (minimum release criteria)	Lot #0021502
Exonuclease Activity (Radioactivity Release) - A 50 μ l reaction in CutSmart TM Buffer containing 1 μ g of a mixture of single and double-stranded [³ H] <i>E. coli</i> DNA and a minimum of 200 units of DraIII-HF TM incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) - After a 10-fold over-digestion of Lambda DNA with DraIII-HF TM , ~75% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with DraIII-HF TM .	Pass
Non-Specific DNase Activity (16 Hour) - A 50 µl reaction in CutSmart [™] Buffer containing 1 µg of Lambda DNA and a minimum of 20 Units of DraIII-HF [™] incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Protein Purity Assay (SDS-PAGE) - DraIII-HF [™] is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	Pass

* The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.

otunen

Authorized by Derek Robinson 24 Apr 2013



-jake

Inspected by Mala Samaranayake 26 Mar 2015