

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: BsiWI-HF®
Catalog Number: R3553S
Concentration: 20,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of PhiX174 DNA in 1 hour at 37°C in a total reaction volume of 50

μl.

Packaging Lot Number: 10060904
Expiration Date: 07/2021
Storage Temperature: -20°C

Storage Conditions: 300 mM NaCl , 10 mM Tris-HCl , 1 mM DTT , 0.1 mM EDTA , 50 %

Glycerol , 500 μg/ml BSA, (pH 7.4 @ 25°C)

Specification Version: PS-R3553S/L v1.0

BsiWI-HF® Component List			
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result
R3553SVIAL	BsiWI-HF®	10049336	Pass
B7204SVIAL	CutSmart® Buffer	10061304	Pass
B7024SVIAL	Gel Loading Dye, Purple (6X)	10055732	Pass

Assay Name/Specification	Lot # 10060904
Protein Purity Assay (SDS-PAGE)	Pass
BsiWI-HF is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	
Non-Specific DNase Activity (16 Hour)	Pass
A 50 µl reaction in CutSmart® Buffer containing 1 µg of PhiX174 DNA and a minimum of	
100 units of BsiWI-HF incubated for 16 hours at 37°C results in a DNA pattern free	
of detectable nuclease degradation as determined by agarose gel electrophoresis.	
Ligation and Recutting (Terminal Integrity)	Pass
After a 20-fold over-digestion of PhiX174 DNA with BsiWI-HF, >95% of the DNA	
fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated	
fragments, >95% can be recut with BsiWI-HF.	
Functional Testing (15 minute Digest)	Pass
A 50 µl reaction in CutSmart® Buffer containing 1 µg of PhiX174 DNA and 1 µl of	
BsiWI-HF incubated for 15 minutes at 37°C results in complete digestion as	



R3553S / Lot: 10060904

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

minimum of 20 units of BsiWI-HF incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.

Stephanie Cornelio Production Scientist

tephoni Oneth

16 Jul 2019

Michael Tonello

Packaging Quality Control Inspector

16 Jan 2020



R3553S / Lot: 10060904

Page 2 of 2