

## New England Biolabs Certificate of Analysis

**Product Name:** BstEII  
**Catalog Number:** R0162M  
**Concentration:** 50,000 U/ml  
**Unit Definition:** One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 60°C in a total reaction volume of 50 µl.  
**Packaging Lot Number:** 10063036  
**Expiration Date:** 01/2022  
**Storage Temperature:** -20°C  
**Storage Conditions:** 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA  
**Specification Version:** PS-R0162T/M v1.0

BstEII Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0162MVIAL	BstEII	10063035	Pass
B7203SVIAL	NEBuffer™ 3.1	10053972	Pass

Assay Name/Specification	Lot # 10063036
<p><b>Endonuclease Activity (Nicking)</b>            A 50 µl reaction in NEBuffer 3.1 containing 1 µg of supercoiled PhiX174 DNA and a minimum of 30 units of BstEII incubated for 4 hours at 60°C results in &lt;10% conversion to the nicked form as determined by agarose gel electrophoresis.</p>	Pass
<p><b>Exonuclease Activity (Radioactivity Release)</b>            A 50 µl reaction in NEBuffer 3.1 containing 1 µg of a mixture of single and double-stranded [<sup>3</sup>H] E. coli DNA and a minimum of 50 units of BstEII incubated for 4 hours at 60°C releases &lt;0.1% of the total radioactivity.</p>	Pass
<p><b>Ligation and Recutting (Terminal Integrity)</b>            After a 10-fold over-digestion of Lambda DNA with BstEII, &gt;95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, &gt;95% can be recut with BstEII.</p>	Pass
<p><b>Non-Specific DNase Activity (16 Hour)</b>            A 50 µl reaction in NEBuffer 3.1 containing 1 µg of Lambda DNA and a minimum of 50 Units of BstEII incubated for 16 hours at 60°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.



---

Stephanie Cornelio  
Production Scientist  
28 Jan 2020



---

Michael Tonello  
Packaging Quality Control Inspector  
28 Jan 2020