

be INSPIRED *drive* DISCOVERY *stay* GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name:	SNAP-Surface® 594
Catalog Number:	S9134S
Packaging Lot Number:	10180561
Expiration Date:	01/2028
Storage Temperature:	-20°C
Specification Version:	PS-S9134S v3.0

SNAP-Surface® 594 Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
S9134SVIAL	SNAP-Surface® 594	10153980	Pass	

Assay Name/Specification	Lot # 10180561
Physical Purity (HPLC)	Pass
SNAP-Surface® 594 is \geq 90% pure as determined by HPLC analysis.	
Identity (Mass Spectrometry)	Pass
The observed molecular mass of SNAP-Surface® 594 is 1057.3 Da +/- 1 Da as determined	
by mass spectrometry analysis.	
In Vitro Protein Labeling	Pass
A 50 μ l reaction in 1X PBS and 1 mM DTT containing 5 μ M of SNAP-tag® Purified	
Protein and a minimum of 10 μ M of SNAP-Surface® 594 is incubated for 1 hour at 37°C	
results in the expected labeled product that is visualized on SDS-PAGE by	
fluorescent detection.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.





be INSPIRED *drive* DISCOVERY *stay* GENUINE

240 County Road Ipswich, MA 01938-2723

Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

Chistopher R. Provort

Chris Provost Production Scientist 20 Jan 2023

Josh Hersey Packaging Quality Control Inspector

09 Feb 2023

