

## New England Biolabs Certificate of Analysis

**Product Name:** *Monarch® DNA Gel Extraction Kit*  
**Catalog Number:** *T1020S*  
**Packaging Lot Number:** *10240878*  
**Expiration Date:** *05/2026*  
**Storage Temperature:** *25°C*  
**Specification Version:** *PS-T1020S/L v2.0*

Monarch® DNA Gel Extraction Kit Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T1034-2	Monarch® DNA Cleanup Columns (5 µg)	10240863	<b>Pass</b>
T1032-2	Monarch® DNA Wash Buffer	10240746	<b>Pass</b>
T1021-2	Monarch® Gel Dissolving Buffer	10240742	<b>Pass</b>
T1018-2	Monarch® Collection Tubes (2 ml)	10240862	<b>Pass</b>
T1016-21	Monarch® DNA Elution Buffer	10240712	<b>Pass</b>

Assay Name/Specification	Lot # 10240878
<b>Functional Testing (DNA Recovery and Purity)</b> Twenty-four Monarch® DNA Cleanup Columns are tested with 1 µg of 1 Kb Ladder resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	<b>Pass</b>
<b>Functional Testing (Gel Purification and Ligation)</b> A linear pLA3 fragment is gel purified and added to 8 Monarch® DNA Cleanup Columns, eluted and ligated with Blunt/TA Ligase Master Mix resulting in ≥90% conversion to a multiband pattern observed by gel electrophoresis in ≥75% of the samples.	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



---

Judith Pacheco  
Production Scientist  
20 Jun 2024



---

Julie Gaumont  
Packaging Quality Control Inspector  
24 Jul 2024