

New England Biolabs Certificate of Analysis

Product Name: *Monarch® RNA Purification Columns*
 Catalog Number: *T2007L*
 Packaging Lot Number: *10166724*
 Expiration Date: *12/2024*
 Storage Temperature: *25°C*
 Specification Version: *PS-T2007L v2.0*

Monarch® RNA Purification Columns Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T2018-1	Monarch® Collection Tubes II	10166717	Pass
T2007-1	Monarch® RNA Purification Columns	10166632	Pass

Assay Name/Specification	Lot # 10166724
Functional Testing (DNA Recovery and Purity) Twenty-four Monarch® RNA Purification Columns are tested with 10 µg of Lambda/BstEII Marker resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Michael Burzyk
Production Scientist
11 Jan 2023



Nikki Samuelson
Packaging Quality Control Inspector
20 Apr 2023