

New England Biolabs Certificate of Analysis

Product Name: Monarch RNA Cleanup Kit (50 µg)
 Catalog Number: T2040L
 Packaging Lot Number: 10109332
 Expiration Date: 11/2022
 Storage Temperature: 25°C
 Specification Version: PS-T2040S/L v2.0

Monarch RNA Cleanup Kit (50 µg) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T2047-1	RNA Cleanup 50 Columns (50 µg) IIC	10086478	Pass
T2042-2	Monarch RNA Cleanup Wash Buffer	10109342	Pass
T2041-2	Monarch RNA Cleanup Binding Buffer	10109341	Pass
T2018-1	Monarch® Collection Tubes II (50)	10086488	Pass
B1500-2	Nuclease-free Water	10109335	Pass


Assay Name/Specification	Lot # 10109332
<p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in Monarch® RNA Cleanup Kit (50 µg) and meet the designated specifications.</p>	Pass
<p>Functional Testing (RNA Recovery and Purity) Twenty-four Monarch® RNA Cleanup Columns (T2047, 50 µg) are tested with 5 µg of RNA resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.</p>	Pass
<p>Functional Testing (Enzyme Inactivation and Removal) Eight Monarch® RNA Cleanup Columns (T2047, 50 µg) are tested with 5 µg of RNA and 2 units of DNase I. A 50 µl reaction in DNase I Reaction Buffer with 1 µg of pBR322 and 50% of the column eluate incubated for 10 minutes at 37°C results in >99.95% removal of DNase I activity in ≥75% of the samples.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Ana Egana
Production Scientist
05 Aug 2021



Michael Burzyk
Packaging Quality Control Inspector
05 Aug 2021