

New England Biolabs Certificate of Analysis

Product Name: *Monarch® Spin Columns S2C and Tubes*
Catalog Number: *T3017L*
Packaging Lot Number: *10250994*
Expiration Date: *07/2026*
Storage Temperature: *25°C*
Specification Version: *PS-T3017S/L v2.0*

| Monarch® Spin Columns S2C and Tubes Component List | | | |
|--|--------------------------------|------------|----------------------|
| NEB Part Number | Component Description | Lot Number | Individual QC Result |
| T3017-1 | Monarch® Spin Columns S2C | 10250217 | Pass |
| T2118-1 | Monarch® Spin Collection Tubes | 10234057 | Pass |


| Assay Name/Specification | Lot # 10250994 |
|---|----------------|
| Functional Testing (DNA Recovery and Purity) Twenty-four Monarch® gDNA Purification Columns are tested with 5 µg of HindIII digested Lambda resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples. | Pass |

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Darcie Spaulding
Production Scientist
08 Aug 2024



Julie Gaumont
Packaging Quality Control Inspector
15 Aug 2024