

## New England Biolabs Certificate of Analysis

Product Name: *Monarch® Spin Columns S2C and Tubes*  
 Catalog Number: *T3017L*  
 Packaging Lot Number: *10250994*  
 Expiration Date: *07/2026*  
 Storage Temperature: *25°C*  
 Specification Version: *PS-T3017S/L v3.0*

Monarch® Spin Columns S2C and Tubes Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T3017-1	Monarch® Spin Columns S2C	10250217	Pass
T2118-1	Monarch® Spin Collection Tubes	10234057	Pass

Assay Name/Specification	Lot # 10250994
<b>Functional Testing (DNA Recovery and Purity)</b> Twenty-four Monarch® Spin Columns S2C are tested with 10 µg of HindIII digested Lambda DNA resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



Darcie Spaulding  
Production Scientist  
08 Aug 2024



Julie Gaumont  
Packaging Quality Control Inspector  
24 Oct 2024