320 Newburyport Turnpike Rowley, MA 01969-2002 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Monarch® HMW DNA Extraction Kit for Tissue

Catalog Number: T3060L
Packaging Lot Number: 10193019
Expiration Date: 05/2025
Storage Temperature: 25°C

Specification Version: PS-T3060S/L v2.0

Monarch® HMW DNA Extraction Kit for Tissue Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T3062-1	Monarch® Protein Separation Solution	10186150	Pass
T3061-1	Monarch® HMW gDNA Tissue Lysis Buffer	10186142	Pass
T3056-1	Monarch® gDNA Elution Buffer II	10195507	Pass
T3018-2	Monarch® RNase A	10184643	Pass
T3015-1	Monarch® gDNA Wash Buffer	10186059	Pass
T3005-1	Monarch® DNA Capture Beads (100)	10186077	Pass
T3004-1	Monarch® Bead Retainers (50)	10095357	Pass
T3003-1	Monarch® 2 ml Tubes (50)	10179464	Pass
T3002-1	Monarch® Pestles (50)	10179457	Pass
T3001-1	Monarch® Pestle Tubes (50)	10179454	Pass
T2018-1	Monarch® Collection Tubes II	10137027	Pass
P8107SVIAL	Proteinase K, Molecular Biology Grade	10198926	Pass

Assay Name/Specification	Lot # 10193019
* Individual Product Component Note	Pass
Standard Quality Control Tests are performed for each component included in Monarch®	
HMW DNA Extraction Kit for Tissue and meet the designated specifications.	
Functional Testing (Genomic DNA Yield and Integrity Analysis) (Tissue)	Pass
Genomic DNA is purified from 8 samples of NEB 10-beta Competent E. coli using the	
High Molecular Weight DNA Extraction from Bacteria protocol. Each set of reagents is	
functionally tested to ensure successful isolation of gDNA from 1 x 10e9 cells.	
Yield of genomic DNA is ≥6 μg in ≥80% of the samples. OD 260/280 and 260/230 are	
≥1.75 in ≥80% of the samples, and DIN values are ≥9 in ≥80% of the samples. In ≥ 80%	
of samples, ≥ 75% of the DNA is 50 kb in size or greater.	

This product has been tested and shown to be in compliance with all specifications.



T3060L / Lot: 10193019 Page 1 of 2



One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Ana Egana Production Scientist

27 Jul 2023

Julie Gaumont

Packaging Quality Control Inspector

27 Jul 2023



T3060L / Lot: 10193019

Page 2 of 2