

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Product Specification

Product Name:	Q5® Hot Start High-Fidelity DNA Polymerase
Catalog #:	M0493S/L
Concentration:	2,000 units/ml
Unit Definition:	One unit is defined as the amount of enzyme that will incorporate 10 nmol of dNTP into acid insoluble material in 30 minutes at 74°C.
Shelf Life:	24 months
Storage Temp:	-20°C
Storage Conditions:	Proprietary
Specification Version:	<i>PS-M0493S/L</i> v1.0
Effective Date:	02 Dec 2015

Assay Name/Specification (minimum release criteria)

Endonuclease Activity (Hot Start, Nicking) - A 50  $\mu$ l reaction in NEBuffer 2 in the presence of 400  $\mu$ M dNTPs containing 1  $\mu$ g of supercoiled pUC19 DNA and a minimum of 10 units of Q5® High-Fidelity DNA Polymerase incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.

PCR Amplification (20 kb Lambda DNA) - A 50  $\mu$ l reaction in Q5® Reaction Buffer in the presence of 200  $\mu$ M dNTPs and 1.0  $\mu$ M primers containing 10 ng Lambda DNA with 1 unit of Q5® Hot Start High-Fidelity DNA Polymerase for 22 cycles of PCR amplification results in the expected 20 kb product.

PCR Amplification (7 kb Human Genomic DNA) - A 50 µl reaction in Q5® Reaction Buffer in the presence of 200 µM dNTPs and 0.5 µM primers containing 20 ng Human Genomic DNA with 1 unit of Q5® Hot Start High-Fidelity DNA Polymerase for 30 cycles of PCR amplification results in the expected 7 kb product.

PCR Amplification (Enhancer Dependent, >65% GC-rich) - A 50  $\mu$ l reaction in Q5® Reaction Buffer and Q5® High GC Enhancer in the presence of 200  $\mu$ M dNTPs and 0.5  $\mu$ M primers containing 20 ng Human Genomic DNA with 1 unit of Q5® Hot Start High-Fidelity DNA Polymerase for 30 cycles of PCR amplification results in the enhancer-dependent production of the expected 452 bp product.

PCR Amplification (Hot Start, Human Genomic DNA) - A 50  $\mu$ l reaction in Q5® Reaction Buffer plus Q5® High GC Enhancer in the presence of 200  $\mu$ M dNTPs and 0.5  $\mu$ M primers containing 100 ng Human Genomic DNA with 1 unit of Q5® Hot Start High-Fidelity DNA Polymerase for 25 cycles of PCR amplification results in the expected 665 bp product, and a decrease in non-specific genomic bands after pre-incubation at room temperature for 1 hour, when compared to a non-hot start control reaction.



PS-M0493S/L v1.0 Page 1 of 2



240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Product Specification

Assay Name/Specification (minimum release criteria)

**Phosphatase Activity (pNPP)** - A 200  $\mu$ l reaction in 1M Diethanolamine, pH 9.8, 0.5 mM MgCl<sub>2</sub> containing 2.5 mM *p*-Nitrophenyl Phosphate (pNPP) and a minimum of 100 units Q5® High-Fidelity DNA Polymerase incubated for 4 hours at 37°C yields <0.0001 unit of alkaline phosphatase activity as determined by spectrophotometric analysis.

**Protein Purity Assay (SDS-PAGE)** - Q5<sup>®</sup> High-Fidelity DNA Polymerase is  $\geq$  95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.

**qPCR DNA Contamination (***E. coli* **Genomic)** - A minimum of 2 units of Q5<sup>®</sup> High-Fidelity DNA Polymerase is screened for the presence of *E. coli* genomic DNA using SYBR<sup>®</sup> Green qPCR with primers specific for the *E. coli* 16S rRNA locus. Results are quantified using a standard curve generated from purified *E. coli* genomic DNA. The measured level of *E. coli* genomic DNA contamination is  $\leq 1 E$ . *coli* genome.

**RNase Activity (Extended Digestion)** - A 10  $\mu$ l reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1  $\mu$ l of Q5® Hot Start High-Fidelity DNA Polymerase is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.

Date 02 Dec 2015

Derek Robinson Director of Quality Control



PS-M0493S/L v1.0 Page 2 of 2