

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: CutSmart® Buffer

Catalog Number: B7204S

Concentration: 10 X Concentrate

Lot Number: 10043025
Expiration Date: 04/2022
Storage Temperature: -20°C

Specification Version: PS-B7204S v1.0

Composition (1X): 50 mM Potassium Acetate , 20 mM Tris Acetate, 10 mM Magnesium

Acetate, 100 μg/ml BSA, (pH 7.9 @ 25°C)

CutSmart® Buffer Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
B7204SVIAL	CutSmart® Buffer	10042783	Pass	

Assay Name/Specification	Lot # 10043025
Conductivity (buffers/solutions)	Pass
The conductivity of 10X CutSmart® Buffer is between 40 and 46 mS at 25°C.	
Endonuclease Activity (Nicking, Buffer) A 50 µl reaction in 1X CutSmart® Buffer containing 1 µg of supercoiled PhiX174 DNA incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Functional Testing (Restriction Digest, Buffer) A 50 μl reaction in 1X CutSmart® Buffer containing 1 μg of Lambda dam- DNA and 1 unit of Clal incubated for 1 hour at 37°C results in complete digestion of the substrate DNA as determined by agarose gel electrophoresis.	Pass
Functional Testing (Restriction Digest, Buffer) A 50 µl reaction in 1X CutSmart® Buffer containing 1 µg of Lambda DNA and 1 unit of Mscl incubated for 1 hour at 37°C results in complete digestion of the substrate DNA as determined by agarose gel electrophoresis.	Pass
Non-Specific DNase Activity (16 hour, Buffer) A 50 µl reaction in 1X CutSmart® Buffer containing 1 µg of PhiX174-HaeIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass



B7204S / Lot: 10043025

Page 1 of 2

substrate RNA remains intact as determined by fluorescent detection.

This product has been tested and shown to be in compliance with all specifications.

Production Scientist

24 Apr 2019

Michael Tonello

Packaging Quality Control Inspector

14 May 2019



B7204S / Lot: 10043025

Page 2 of 2