

New England Biolabs Certificate of Analysis

Product Name: NEB[®] PCR Cloning Kit (Without Competent Cells)
 Catalog Number: E1203S
 Packaging Lot Number: 10171295
 Expiration Date: 08/2023
 Storage Temperature: -20°C
 Specification Version: PS-E1203S v1.0

NEB [®] PCR Cloning Kit (Without Competent Cells) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
S1513AVIAL	Cloning Analysis Reverse Primer	10155091	Pass
S1512AVIAL	Cloning Analysis Forward Primer	10155090	Pass
N0555AVIAL	Amplicon Cloning Control	10155095	Pass
N0312AVIAL	Linearized pMiniT [™] 2.0 Vector	10155094	Pass
M1016AVIAL	Cloning Master Mix 2	10155093	Pass
M1015AVIAL	Cloning Master Mix 1	10155092	Pass

Assay Name/Specification	Lot # 10171295
<p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEB[®] PCR Cloning Kit (Without Competent Cells) and meet the designated specifications.</p> <p>Functional Testing (NEB[®] PCR Cloning Kit) A 10 µl ligation reaction containing 5 µl cloning mix, 25 ng of linearized pMiniT[™] 2.0 vector and 30 ng of a 1 kb amplicon containing A overhangs is incubated for 5 minutes at 25°C. Transformation of NEB 10-beta Competent E. coli with 2 µl of the ligated master mix/fragment mixture yields ≥ 100 colonies when 5% of the transformation is plated on an ampicillin plate and incubated overnight at 37°C. The colonies are screened by colony PCR and ≥95% of colonies have an insert present.</p>	<p>Pass</p> <p>Pass</p>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Cathy Shea

Cathy Shea
Production Scientist
17 Aug 2022

Michael Tonello

Michael Tonello
Packaging Quality Control Inspector
19 Dec 2022