

www.neb.com info@neb.com



New England Biolabs Certificate of Analysis

Product Name: Gibson Assembly® Master Mix

Catalog Number: E2611L
Packaging Lot Number: 10236994
Expiration Date: 02/2025
Storage Temperature: -20°C

Specification Version: PS-E2611S/L v2.0

| Gibson Assembly® Master Mix Component List | | | | |
|--|-----------------------------|------------|----------------------|--|
| NEB Part Number | Component Description | Lot Number | Individual QC Result | |
| N2611AVIAL | NEBuilder® Positive Control | 10233980 | Pass | |
| M5510AAVIAL | Gibson Assembly® Master Mix | 10230124 | Pass | |

| Assay Name/Specification | Lot # 10236994 |
|--|----------------|
| * Individual Product Component Note Standard Quality Control Tests are performed for each component included in Gibson | Pass |
| Assembly® Master Mix and meet the designated specifications. | |
| Functional Testing (Gibson Assembly) | Pass |
| A 20 µl reaction containing 10 µl of 2X Gibson Assembly Master Mix and six 0.05 pmol | |
| fragments of pUC19 (five 400 bp fragments and one 2,780 bp fragment, each with a 40 | |
| bp overlap) is incubated at 50°C for 60 minutes. Transformation of NEB 5-alpha competent E. coli cells (NEB #C2987) with 2 µl of the master mix/fragment mixture | |
| yields greater than 100 white colonies on an ampicillin plate with IPTG/X-Gal after | |
| overnight incubation at 37°C. | |

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



E2611L / Lot: 10236994

Page 1 of 2



Ana Egana Production Scientist 08 May 2024 Michael Tonello

Packaging Quality Control Inspector

08 May 2024