

## New England Biolabs Certificate of Analysis


Product Name: Luna® SARS-CoV-2 RT-qPCR Multiplex Assay Kit  
 Catalog Number: E3019S  
 Packaging Lot Number: 10166257  
 Expiration Date: 08/2023  
 Storage Temperature: -20°C  
 Specification Version: PS-E3019S/L v1.0

Luna® SARS-CoV-2 RT-qPCR Multiplex Assay Kit Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
S0631AVIAL	SARS-CoV-2 Primer/Probe Mix (N1/N2/RP)	10149318	Pass
N2117SVIAL	SARS-CoV-2 Positive Control (N gene)	10143730	Pass
M3019SVIAL	Luna® Probe One-Step RT-qPCR 4X Mix with UDG	10162006	Pass
B1502AVIAL	Nuclease-free Water	10154305	Pass

Assay Name/Specification	Lot # 10166257
<p><b>Functional Testing (RT-qPCR, SARS-CoV-2)</b>            Luna® SARS-CoV-2 RT-qPCR Multiplex Assay Kit is functionally tested and compared to a control lot using 10 ng of human total RNA spiked with synthetic SARS-CoV-2 RNA in an RT-qPCR assay, resulting in a standard curve with a calculated qPCR efficiency of 90-110%, and a dynamic range of 5 orders of magnitude.</p>	Pass
<p><b>* Individual Product Component Note</b>            Standard Quality Control Tests are performed for each component included in Luna® SARS-CoV-2 RT-qPCR Multiplex Assay Kit and meet the designated specifications.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.

  
Lea Antonopoulos  
Production Scientist  
31 Aug 2022

  
Michael Tonello  
Packaging Quality Control Inspector  
19 Dec 2022