

## New England Biolabs Certificate of Analysis

**Product Name:** NEBNext<sup>®</sup> Multiplex Oligos for Illumina<sup>®</sup> (96 Unique Dual Index Primer Pairs Set 4)  
**Catalog Number:** E6446S  
**Packaging Lot Number:** 10076753  
**Expiration Date:** 05/2022  
**Storage Temperature:** -20°C  
**Specification Version:** PS-E6446S/L v1.0

NEBNext <sup>®</sup> Multiplex Oligos for Illumina <sup>®</sup> (96 Unique Dual Index Primer Pairs Set 4) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E6612AVIAL	NEBNext <sup>®</sup> Adaptor for Illumina <sup>®</sup>	10075114	Pass
E6610AVIAL	USER Enzyme	10075115	Pass
E6447AVIAL	NEBNext <sup>®</sup> Unique Dual Index Primer Set 4	10075116	Pass

Assay Name/Specification	Lot # 10076753
<p><b>Functional Testing (Library Construction, Oligo)</b>            Each of the components are functionally validated through construction of libraries made from commercially available genomic DNA. Libraries are made from each unique barcoded oligonucleotide and are sequenced together on the same Illumina<sup>®</sup> flow cell and compared across various metrics including library yield, fraction of reads aligning to the reference, GC bias, and insert size.</p>	Pass
<p><b>* Individual Product Component Note</b>            Standard Quality Control Tests are performed for each component included in NEBNext<sup>®</sup> Multiplex Oligos for Illumina<sup>®</sup> (96 Unique Dual Index Primer Pairs Set 4) and meet the designated specifications.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.

*Christine Sumner*

---

Christine Sumner  
Production Scientist  
05 Jun 2020

*Michael Tonello*

---

Michael Tonello  
Packaging Quality Control Inspector  
05 Jun 2020