

New England Biolabs Certificate of Analysis

Product Name: NEBNext® Multiplex Oligos for Illumina® (Index Primers Set 1)
Catalog Number: E7335S
Packaging Lot Number: 10172559
Expiration Date: 01/2025
Storage Temperature: -20°C
Specification Version: PS-E7335S/L v1.0

NEBNext® Multiplex Oligos for Illumina® (Index Primers Set 1) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7338AVIAL	USER® Enzyme	10172561	Pass
E7337AVIAL	NEBNext® Adaptor for Illumina®	10172560	Pass
E7322AVIAL	NEBNext® Index 12 Primer for Illumina®	10172580	Pass
E7321AVIAL	NEBNext® Index 11 Primer for Illumina®	10172578	Pass
E7320AVIAL	NEBNext® Index 10 Primer for Illumina®	10172576	Pass
E7319AVIAL	NEBNext® Index 9 Primer for Illumina®	10172574	Pass
E7318AVIAL	NEBNext® Index 8 Primer for Illumina®	10172572	Pass
E7317AVIAL	NEBNext® Index 7 Primer for Illumina®	10172571	Pass
E7316AVIAL	NEBNext® Index 6 Primer for Illumina®	10172569	Pass
E7315AVIAL	NEBNext® Index 5 Primer for Illumina®	10172568	Pass
E7314AVIAL	NEBNext® Index 4 Primer for Illumina®	10172567	Pass
E7313AVIAL	NEBNext® Index 3 Primer for Illumina®	10172566	Pass
E7312AVIAL	NEBNext® Index 2 Primer for Illumina®	10172565	Pass
E7311AVIAL	NEBNext® Index 1 Primer for Illumina®	10172563	Pass
E6861AVIAL	NEBNext® Universal PCR Primer for Illumina®	10172562	Pass

Assay Name/Specification	Lot # 10172559
<p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Multiplex Oligos for Illumina® (Index Primers Set 1) and meet the designated specifications.</p>	Pass
<p>Functional Testing (Library Construction, Oligo) Each set of reagents is functionally validated through construction of libraries made from commercially available genomic DNA. Libraries are made from each unique barcoded oligonucleotide and are sequenced together on the same Illumina flow cell and compared across various metrics including library yield, fraction of reads</p>	Pass

Assay Name/Specification	Lot # 10172559
aligning to the reference, GC bias, and insert size.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Christine Sumner
Production Scientist
21 Feb 2023



Josh Hersey
Packaging Quality Control Inspector
25 Aug 2023