

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: NEBNext® Ultra™ II DNA Library Prep Kit for Illumina®

Catalog Number: E7645L
Lot Number: 10045242
Expiration Date: 03/2020
Storage Temperature: -20°C

Specification Version: PS-E7645S/L v1.0

| NEBNext® Ultra™ II DNA Library Prep Kit for Illumina® Component List | | | | |
|--|---|------------|----------------------|--|
| NEB Part Number | Component Description | Lot Number | Individual QC Result | |
| E7649AAVIAL | NEBNext® Ultra™ II Q5® Master Mix | 10038053 | Pass | |
| E7648AAVIAL | NEBNext® Ultra™ II Ligation Master Mix | 10038051 | Pass | |
| E7647AAVIAL | NEBNext® Ultra™ II End Prep Reaction Buffer | 10038050 | Pass | |
| E7646AAVIAL | NEBNext® Ultra™ II End Prep Enzyme Mix | 10038049 | Pass | |
| E7374AAVIAL | NEBNext® Ligation Enhancer | 10038052 | Pass | |

| Assay Name/Specification | Lot # 10045242 |
|--|----------------|
| Functional Testing (Library Construction, DNA) | Pass |
| Each set of reagents is functionally validated and compared to a previous lot | |
| through construction of libraries made from commercially available genomic DNA, | |
| using the kit's minimum and maximum input requirements. Libraries made from previous | |
| and current lots for both input DNA amounts are sequenced together on the same | |
| Illumina® flow cell and compared across various metrics including library yield, | |
| fraction of reads aligning to the reference, GC bias, and insert size. | |
| * Individual Product Component Note | Pass |
| Standard Quality Control Tests are performed for each component included in NEBNext® | |
| Ultra™ II DNA Library Prep Kit for Illumina® and meet the designated specifications. | |

This product has been tested and shown to be in compliance with all specifications.



E7645L / Lot: 10045242

Page 1 of 2

Christin Summ

Christine Sumner Production Scientist 29 May 2019 Michael Tonello

Packaging Quality Control Inspector

29 May 2019

Page 2 of 2