

## New England Biolabs Certificate of Analysis

Product Name: NEBNext<sup>®</sup> Ultra<sup>™</sup> II DNA Library Prep Kit for Illumina<sup>®</sup>  
 Catalog Number: E7645S  
 Packaging Lot Number: 10064315  
 Expiration Date: 12/2020  
 Storage Temperature: -20°C  
 Specification Version: PS-E7645S/L v1.0

NEBNext <sup>®</sup> Ultra <sup>™</sup> II DNA Library Prep Kit for Illumina <sup>®</sup> Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7649AVIAL	NEBNext <sup>®</sup> Ultra <sup>™</sup> II Q5 <sup>®</sup> Master Mix	10059702	Pass
E7648AVIAL	NEBNext <sup>®</sup> Ultra <sup>™</sup> II Ligation Master Mix	10059700	Pass
E7647AVIAL	NEBNext <sup>®</sup> Ultra <sup>™</sup> II End Prep Reaction Buffer	10059699	Pass
E7646AVIAL	NEBNext <sup>®</sup> Ultra <sup>™</sup> II End Prep Enzyme Mix	10059698	Pass
E7374AVIAL	NEBNext <sup>®</sup> Ligation Enhancer	10059701	Pass

Assay Name/Specification	Lot # 10064315
<b>Functional Testing (Library Construction, DNA)</b> Each set of reagents is functionally validated and compared to a previous lot through construction of libraries made from commercially available genomic DNA, using the kit's minimum and maximum input requirements. Libraries made from previous and current lots for both input DNA amounts are sequenced together on the same Illumina <sup>®</sup> flow cell and compared across various metrics including library yield, fraction of reads aligning to the reference, GC bias, and insert size.	Pass
<b>* Individual Product Component Note</b> Standard Quality Control Tests are performed for each component included in NEBNext <sup>®</sup> Ultra <sup>™</sup> II DNA Library Prep Kit for Illumina <sup>®</sup> and meet the designated specifications.	Pass

This product has been tested and shown to be in compliance with all specifications.

*Christine Sumner*

---

Christine Sumner  
Production Scientist  
07 Jan 2020

*Michael Tonello*

---

Michael Tonello  
Packaging Quality Control Inspector  
07 Jan 2020