

New England Biolabs Certificate of Analysis

Product Name: NEBNext® Ultra™ II FS DNA Module
Catalog Number: E7810L
Packaging Lot Number: 10231629
Expiration Date: 02/2025
Storage Temperature: -20°C
Specification Version: PS-E7810S/L v1.0

NEBNext® Ultra™ II FS DNA Module Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7808AAVIAL	TE Buffer	10231635	Pass
E7807AAVIAL	NEBNext® Ultra™ II FS Reaction Buffer	10231634	Pass
E7806AAVIAL	NEBNext® Ultra™ II FS Enzyme Mix	10231633	Pass

Assay Name/Specification	Lot # 10231629
<p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Ultra™ II FS DNA Module and meet the designated specifications.</p>	Pass
<p>Functional Testing (Library Construction, FS DNA) Each set of reagents is functionally validated and compared to the previous lot through construction of libraries made from commercially available genomic DNA, using the kit's minimum and maximum input requirements. A fragmentation time of 20 minutes was used to generate an insert size of approximately 200 bp. The final average library size is between 270 and 450 bp as determined by an Agilent Bioanalyzer. Libraries made from the previous and current lots for both input DNA amounts are sequenced together on the same Illumina flow cell and compared across various metrics including library yield, fraction of reads aligning to the reference, GC bias, and insert size.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Christine Sumner
Production Scientist
08 Apr 2024



Josh Hersey
Packaging Quality Control Inspector
10 May 2024