

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Cas9 Nuclease, S. pyogenes

Catalog Number: M0386M
Concentration: 20 µM
Packaging Lot Number: 10091235
Expiration Date: 09/2022
Storage Temperature: -20°C

Storage Conditions: 10 mM Tris-HCl, 300 mM NaCl, 1 mM DTT, 0.1 mM EDTA, 50 %

Glycerol, (pH 7.4 @ 25°C)

Specification Version: PS-M0386T/M v1.0

Cas9 Nuclease, S. pyogenes Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
M0386MVIAL	Cas9 Nuclease, S. pyogenes	10084276	Pass	
B7203SVIAL	NEBuffer™ 3.1	10085493	Pass	

Assay Name/Specification	Lot # 10091235
Endonuclease Activity (Nicking) A 50 μl reaction in NEBuffer 3.1 containing 1 μg of supercoiled PhiX174 DNA and a minimum of 1 pmol of Cas9 Nuclease, S. pyogenes incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 1 pmol of Cas9 Nuclease, S. pyogenes incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Functional Testing (Targeted Digestion) A 20 µl reaction in NEBuffer 3.1 containing 20 nM of 100 bp FAM and ROX-labeled double-stranded target DNA, 100 nM sgRNA, and 100 nM Cas9 Nuclease, S. pyogenes incubated for 1 hour at 37°C results in ≥90% targeted digestion of the substrate DNA as determined by capillary electrophoresis.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of Lambda DNA and a minimum of 1 pmol of Cas9 Nuclease, S. pyogenes incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel	Pass



M0386M / Lot: 10091235

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

gel electrophoresis using fluorescent detection.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Bhairavi Jani Production Scientist 22 Nov 2020 Michael Tonello

Packaging Quality Control Inspector

22 Nov 2020



M0386M / Lot: 10091235

Page 2 of 2