


## New England Biolabs Certificate of Analysis

Product Name: Luna<sup>®</sup> Universal qPCR Master Mix  
 Catalog Number: M3003L  
 Concentration: 2 X Concentrate  
 Lot Number: 10015440  
 Expiration Date: 03/2020  
 Storage Temperature: -20°C  
 Specification Version: PS-M3003S/L/G/X/E v1.0  
 Composition (1X): Proprietary

Luna <sup>®</sup> Universal qPCR Master Mix Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
M3003SVIAL	Luna <sup>®</sup> Universal qPCR Master Mix	0101803	Pass

Assay Name/Specification	Lot # 10015440
<p><b>Functional Testing (qPCR)</b>            Luna<sup>®</sup> Universal qPCR Master Mix is functionally tested in qPCR with human cDNA template, resulting in a standard curve with a calculated qPCR efficiency of 90-110%, and a dynamic range of 5 orders of magnitude.</p>	Pass
<p><b>Non-Specific DNase Activity (16 hour, Master Mix)</b>            A 50 µl reaction in 1X Luna<sup>®</sup> Universal qPCR Master Mix containing 1 µg of T3 DNA in addition to a reaction containing Lambda-HindIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.</p>	Pass
<p><b>qPCR DNA Contamination (E. coli Genomic)</b>            A minimum of 1 µl of Luna<sup>®</sup> Universal qPCR Master Mix is screened for the presence of E. coli genomic DNA using SYBR<sup>®</sup> Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome.</p>	Pass
<p><b>RNase Activity Assay (4 Hour Digestion)</b>            A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 µl of Luna<sup>®</sup> Universal qPCR Master Mix is incubated at 37°C. After incubation for 4 hours, &gt;90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.



---

Lynne Apone  
Production Scientist  
10 Jul 2018



---

Josh Hersey  
Packaging Quality Control Inspector  
10 Jul 2018