

be INSPIRED *drive* DISCOVERY *stay* GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name:	Ribonucleotide Solution Mix
Catalog Number:	N0466S
Concentration:	25 mM
Packaging Lot Number:	10140299
Expiration Date:	08/2023
Storage Temperature:	-20°C
Storage Conditions:	Supplied as Sodium salts, pH 7.5
Specification Version:	PS-N0466S/L v1.0

Ribonucleotide Solution Mix Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
N0466SVIAL	Ribonucleotide Solution Mix	10118138	Pass	

Assay Name/Specification	Lot # 10140299
Functional Testing (High Yield RNA Synthesis)	Pass
A 20 µl reaction in transcription reaction buffer containing 1 µg FLuc Control Template, 2 µl T7 RNA Polymerase Mix, and 10 mM ribonucleotides (each) incubated for	
2 hours at 37°C yields \geq 150 µg RNA transcript.	
Endonuclease Activity (Nicking)	Pass
A 50 µl reaction in NEBuffer 4 containing 1 µg of supercoiled PhiX174 DNA and a minimum of 500 nmol of Ribonucleotide Solution Mix incubated for 16 hours at 37°C	
results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	
RNase Activity (Extended Digestion)	Pass
A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA	
and a minimum of 100 nmol of Ribonucleotide Solution Mix is incubated at 37°C. After	
incubation for 16 hours, >90% of the substrate RNA remains intact as determined by	
gel electrophoresis using gel electrophoresis.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.





be INSPIRED drive DISCOVERY stay GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

Bhairavi Jani Production Scientist 31 Mar 2022

Michae 711.

Michael Tonello Packaging Quality Control Inspector 31 Mar 2022

