

New England Biolabs Certificate of Analysis

Product Name: *Lambda DNA*
Catalog #: *N3011S/L*
Concentration: *500 µg/ml*
Unit Definition: *N/A*
Lot #: *1661710*
Assay Date: *10/2017*
Expiration Date: *10/2019*
Storage Temp: *-20°C*
Storage Conditions: *10 mM Tris-HCl (pH 8.0), 1 mM EDTA*
Specification Version: *PS-N3011S/L v2.0*
Effective Date: *03 May 2016*

Assay Name/Specification (minimum release criteria)	Lot #1661710
A260/A280 Assay - The ratio of UV absorption of Lambda DNA at 260 and 280 nm is between 1.8 and 2.0.	Pass
DNA Concentration (A260) - The concentration of Lambda DNA is between 500 and 550 µg/ml as determined by UV absorption at 260 nm.	Pass
Electrophoretic Pattern (Linear DNA) - The banding pattern of Lambda DNA on a 1.2% agarose gel is evaluated against a control lot for sharpness and relative intensity as determined by gel electrophoresis using Ethidium Bromide.	Pass
Non-Specific DNase Activity (DNA, 16 hour) - A 50 µl reaction in 1X NEBuffer 2 containing 2.5 µg of Lambda DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Restriction Digest (Correct Pattern) - A 50 µl reaction in NEBuffer 2.1 containing 2.5 µg of Lambda DNA and 20 units of HindIII incubated for 1 hour at 37°C produces the expected pattern of DNA fragments as determined by agarose gel electrophoresis.	Pass



Authorized by
Derek Robinson
03 May 2016



Inspected by
Vanessa Mathieu-Sheltry
13 Oct 2017

