

New England Biolabs Certificate of Analysis

Product Name: *PhiX174 RF II DNA*
Catalog Number: *N3022L*
Concentration: *1,000 µg/ml*
Unit Definition: *N/A*
Packaging Lot Number: *10058580*
Expiration Date: *11/2021*
Storage Temperature: *-20°C*
Storage Conditions: *10 mM Tris-HCl (pH 8.0), 1 mM EDTA*
Specification Version: *PS-N3022S/L v1.0*

PhiX174 RF II DNA Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
N3022LVIAL	PhiX174 RF II DNA	10058579	Pass

Assay Name/Specification	Lot # 10058580
Electrophoretic Pattern (Plasmid) The banding pattern of ϕ X174 RF II DNA on a 1.2% agarose gel is evaluated against a control lot for sharpness and relative intensity as determined by gel electrophoresis using Ethidium Bromide.	Pass
Non-Specific DNase Activity (DNA, 16 hour) A 50 µl reaction in 1X NEBuffer 2 containing 5 µg of ϕ X174 RF II DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Restriction Digest (Linearization) A 50 µl reaction in CutSmart™ Buffer containing 5 µg of ϕ X174 RF II DNA DNA and 20 units of XhoI incubated for 1 hour at 37°C produces > 95% linearization resulting in a single band of approximately 5386 bp as determined by agarose gel electrophoresis.	Pass
A260/A280 Assay The ratio of UV absorption of ϕ X174 RF II DNA at 260 and 280 nm is between 1.8 and 2.0.	Pass
DNA Concentration (A260) The concentration of ϕ X174 RF II DNA is between 1000 and 1050 µg/ml as determined by UV absorption at 260 nm.	Pass

This product has been tested and shown to be in compliance with all specifications.



Vanessa Mathieu-Sheltry
Production Scientist
01 Nov 2019



Michael Tonello
Packaging Quality Control Inspector
21 Nov 2019