

New England Biolabs Certificate of Analysis

Product Name: HpaI
Catalog Number: R0105S
Concentration: 5,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in rCutSmart Buffer in 1 hour at 37°C in a total reaction volume of 50 µl.
Packaging Lot Number: 10181023
Expiration Date: 12/2024
Storage Temperature: -20°C
Storage Conditions: 10 mM Tris-HCl, 50 mM KCl, 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml rAlbumin (pH 7.4 @25°C)
Specification Version: PS-R0105S/L/V v2.0


HpaI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0105SVIAL	HpaI	10173215	Pass
B7024AVIAL	Gel Loading Dye, Purple (6X)	10175294	Pass
B6004SVIAL	rCutSmart™ Buffer	10176423	Pass

Assay Name/Specification	Lot # 10181023
Endonuclease Activity (Nicking) A 50 µl reaction in rCutSmart™ Buffer containing 1 µg of supercoiled pUC19 DNA and a minimum of 15 units of HpaI incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in rCutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 50 units of HpaI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of Lambda DNA with HpaI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with HpaI.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in rCutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of	Pass


Assay Name/Specification	Lot # 10181023
<p>15 units of HpaI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.</p>	
<p>Protein Purity Assay (SDS-PAGE) HpaI is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.</p>	Pass
<p>qPCR DNA Contamination (E. coli Genomic) A minimum of 5 units of HpaI is screened for the presence of E. coli genomic DNA using SYBR® Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



YunJie Sun
Production Scientist
06 Dec 2022



Josh Hersey
Packaging Quality Control Inspector
23 Mar 2023