

## New England Biolabs Certificate of Analysis

**Product Name:** BstEII  
**Catalog Number:** R0162L  
**Concentration:** 10,000 U/ml  
**Unit Definition:** One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 60°C in a total reaction volume of 50 µl.  
**Lot Number:** 10008590  
**Expiration Date:** 04/2020  
**Storage Temperature:** -20°C  
**Storage Conditions:** 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA  
**Specification Version:** PS-R0162S/L v1.0

BstEII Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0162LVIAL	BstEII	0851804	Pass
B7203SVIAL	NEBuffer™ 3.1	0541804	Pass

Assay Name/Specification	Lot # 10008590
<b>Endonuclease Activity (Nicking)</b> A 50 µl reaction in NEBuffer 3.1 containing 1 µg of supercoiled PhiX174 DNA and a minimum of 30 units of BstEII incubated for 4 hours at 60°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
<b>Non-Specific DNase Activity (16 Hour)</b> A 50 µl reaction in NEBuffer 3.1 containing 1 µg of Lambda DNA and a minimum of 50 Units of BstEII incubated for 16 hours at 60°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
<b>Exonuclease Activity (Radioactivity Release)</b> A 50 µl reaction in NEBuffer 3.1 containing 1 µg of a mixture of single and double-stranded [ <sup>3</sup> H] E. coli DNA and a minimum of 50 units of BstEII incubated for 4 hours at 60°C releases <0.1% of the total radioactivity.	Pass
<b>Ligation and Recutting (Terminal Integrity)</b> After a 10-fold over-digestion of Lambda DNA with BstEII, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with BstEII.	Pass

This product has been tested and shown to be in compliance with all specifications.



Penghua Zhang  
Production Scientist  
26 Jun 2018



Michael Tonello  
Packaging Quality Control Inspector  
26 Jun 2018