

New England Biolabs Certificate of Analysis

Product Name: Xmal
Catalog Number: R0180S
Concentration: 10,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg pXba in 1 hour at 37°C in a total reaction volume of 50 µl.
Lot Number: 10043451
Expiration Date: 02/2021
Storage Temperature: -20°C
Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA
Specification Version: PS-R0180S/L v1.0

Xmal Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0180SVIAL	Xmal	10035142	Pass
B7204SVIAL	CutSmart® Buffer	10042964	Pass
B7024SVIAL	Gel Loading Dye, Purple (6X)	10038710	Pass

Assay Name/Specification	Lot # 10043451
Endonuclease Activity (Nicking) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 50 Units of Xmal incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 50 units of Xmal incubated for 4 hours at 37°C releases <0.2% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of pXba DNA with Xmal, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Xmal.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of pXba DNA and a minimum of 10 Units of Xmal incubated for 16 hours at 37°C results in a DNA pattern free of	Pass

Assay Name/Specification	Lot # 10043451
<p>detectable nuclease degradation as determined by agarose gel electrophoresis.</p> <p>Protein Purity Assay (SDS-PAGE) Xmal is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.</p>	<p>Pass</p>

This product has been tested and shown to be in compliance with all specifications.



Penghua Zhang
Production Scientist
30 Jan 2019



Michael Tonello
Packaging Quality Control Inspector
08 May 2019