

New England Biolabs Certificate of Analysis

Product Name: Ascl
Catalog Number: R0558S
Concentration: 10,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.
Lot Number: 10019462
Expiration Date: 01/2020
Storage Temperature: -20°C
Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA
Specification Version: PS-R0558S/L v1.0

Ascl Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0558SVIAL	Ascl	0131801	Pass
B7204SVIAL	CutSmart® Buffer	10014372	Pass
B7024SVIAL	Gel Loading Dye, Purple (6X)	10013725	Pass

Assay Name/Specification	Lot # 10019462
Blue-White Screening (Terminal Integrity) A sample of pNEB193 vector linearized with a 10-fold excess of Ascl, religated and transformed into an E. coli strain expressing the LacZ beta fragment gene results in <1% white colonies.	Pass
Endonuclease Activity (Nicking) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 30 Units of Ascl incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 100 units of Ascl incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of Lambda DNA with Ascl, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments,	Pass

Assay Name/Specification	Lot # 10019462
<p>>95% can be recut with Ascl.</p> <p>Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 30 Units of Ascl incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.</p>	<p>Pass</p>

This product has been tested and shown to be in compliance with all specifications.



Tony Spear-Alfonso
Production Scientist
10 Jul 2018



Michael Tonello
Packaging Quality Control Inspector
15 Aug 2018