


## New England Biolabs Certificate of Analysis

**Product Name:** Bpml  
**Catalog Number:** R0565S  
**Concentration:** 2,000 U/ml  
**Unit Definition:** One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.  
**Lot Number:** 10011698  
**Expiration Date:** 06/2020  
**Storage Temperature:** -20°C  
**Storage Conditions:** 200 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA  
**Specification Version:** PS-R0565S/L v2.0

Bpml Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0565SVIAL	Bpml	10011699	Pass
B7203SVIAL	NEBuffer™ 3.1	10010189	Pass

Assay Name/Specification	Lot # 10011698
<p><b>Non-Specific DNase Activity (16 hour)</b>            A 50 µl reaction in NEBuffer 3.1 containing 1 µg of Lambda DNA and a minimum of 2 Units of Bpml incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. NOTE: although no nuclease degradation is detected under these conditions, extended incubations and/or high concentrations of this enzyme may result in star activity. See the product FAQ for recommended reaction conditions for this enzyme.</p>	Pass
<p><b>Exonuclease Activity (Radioactivity Release)</b>            A 50 µl reaction in NEBuffer 3.1 containing 1 µg of a mixture of single and double-stranded [<sup>3</sup>H] E. coli DNA and a minimum of 10 units of Bpml incubated for 4 hours at 37°C releases &lt;0.1% of the total radioactivity.</p>	Pass
<p><b>Ligation and Recutting (Terminal Integrity)</b>            After a 10-fold over-digestion of Lambda DNA with Bpml, &gt;95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, ~75% can be recut with Bpml.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.



Tony Spear-Alfonso  
Production Scientist  
06 Jun 2018



Josh Hersey  
Packaging Quality Control Inspector  
03 Jul 2018