

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: Protein G Magnetic Beads

Catalog Number: \$1430\$
Packaging Lot Number: 10077450
Expiration Date: 09/2022
Storage Temperature: 4°C

Storage Conditions: 0.02% NaN3, 0.1 % BSA, 0.05 % Tween® 20, 1 X PBS, (pH 7.4 @ 25°C)

Specification Version: PS-S1430S v2.0

Protein G Magnetic Beads Component List			
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result
S1430SVIAL	Protein G Magnetic Beads	10052588	Pass

Assay Name/Specification	Lot # 10077450
RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 µl of Protein G Magnetic Beads is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	Pass
Binding Capacity (Magnetic Beads) Protein G Magnetic Beads ( 100 μl ) were equilibrated and incubated with 500 μl of Human Serum IgG for 1 hour at 25°C, then washed and the IgG eluted. Binding capacity was determined to be >280 μg of IgG per ml of beads.	Pass
Functional Binding Assay (Qualitative) Protein G Magnetic Beads (100 μl) were equilibrated and incubated with 500 μl of Human Serum IgG for 1 hour at 25°C, then washed, eluted and evaluated by Tris-Glycine gel to confirm low non-specific binding of extract proteins and high isolation of target.	Pass
Non-Specific DNase Activity (16 hour, Buffer) A 50 µl reaction in Protein G Magnetic Bead Storage Buffer containing 1 µg of PhiX174-HaeIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit



S1430S / Lot: 10077450

Page 1 of 2

www.neb.com/trademarks for additional information.

Production Scientist

16 Jul 2020

Josh Hersey Packaging Quality Control Inspector

16 Jul 2020



S1430S / Lot: 10077450

Page 2 of 2