


## New England Biolabs Certificate of Analysis

**Product Name:** Monarch Plasmid Miniprep Columns  
**Catalog Number:** T1017L  
**Concentration:** 1 X Concentrate  
**Packaging Lot Number:** 10040690  
**Expiration Date:** 04/2022  
**Storage Temperature:** 25°C  
**Specification Version:** PS-T1017L v1.0

Assay Name/Specification	Lot # 10040690
<b>Functional Testing (DNA Recovery and Purity)</b> Twenty-four Monarch® Plasmid Miniprep Columns are tested with 1.5 µg of 1 Kb Ladder resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.



Ana Egana  
 Production Scientist  
 08 Apr 2020

Darcie Spaulding  
 Packaging Quality Control Inspector  
 08 Apr 2020