


New England Biolabs Certificate of Analysis

Product Name: *Monarch DNA Gel Extraction Kit*
Catalog Number: *T1020S*
Lot Number: *0101811*
Expiration Date: *11/2020*
Storage Temperature: *25°C*
Specification Version: *PS-T1020S/L v1.0*

Assay Name/Specification	Lot # 0101811
<p>Functional Testing (DNA Recovery and Purity) Twenty-four Monarch® DNA Cleanup Columns are tested with 1 µg of 1 Kb Ladder resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.</p>	Pass
<p>Functional Testing (Gel Purification and Ligation) A linear pLA3 fragment is gel purified and added to 8 Monarch® DNA Cleanup Columns, eluted and ligated with Blunt/TA Ligase Master Mix resulting in ≥90% conversion to a multiband pattern observed by gel electrophoresis in ≥75% of the samples.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.



Ana Egana
 Production Scientist
 29 Oct 2018



Heidi Church
 Packaging Quality Control Inspector
 29 Oct 2018