

New England Biolabs Certificate of Analysis


Product Name: Monarch® Spin DNA Gel Extraction Kit
Catalog Number: T1120G
Packaging Lot Number: 10276557
Expiration Date: 02/2027
Storage Temperature: 25°C
Specification Version: PS-T1120S/L/G v1.0

Monarch® Spin DNA Gel Extraction Kit Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T2118-21	Monarch® Spin Collection Tubes	10276349	Pass
T2037-21	Monarch® Spin Columns S1A	10276076	Pass
T1121-21	Monarch® Buffer BY	10276288	Pass
T1116-21	Monarch® Buffer EY	10276284	Pass
T1115-21	Monarch® Buffer WZ	10276281	Pass

Assay Name/Specification	Lot # 10276557
Functional Testing (DNA Recovery and Purity) Twenty-four Monarch® Spin Columns S1A are tested with 1 µg of 1 kb DNA Ladder resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	Pass
Functional Testing (Gel Purification and Ligation) A linear pLA3 fragment is gel purified and added to 8 Monarch® Spin Columns S1A, eluted and ligated with Blunt/TA Ligase Master Mix resulting in ≥90% conversion to a multiband pattern observed by gel electrophoresis in ≥75% of the samples.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Nikki Samuelson
Production Scientist
04 Mar 2025



Julie Gaumont
Packaging Quality Control Inspector
10 Apr 2025