


New England Biolabs Certificate of Analysis

Product Name: Monarch RNA Purification Columns
Catalog Number: T2007L
Packaging Lot Number: 10079497
Expiration Date: 03/2022
Storage Temperature: 25°C
Specification Version: PS-T2007L v1.0

Assay Name/Specification	Lot # 10079497
Functional Testing (DNA Recovery and Purity) Twenty-four Monarch [®] RNA Purification Columns are tested with 10 µg of Lambda/BstEII Marker resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Ana Egana
 Production Scientist
 26 Jan 2021



Heidi Church
 Packaging Quality Control Inspector
 26 Jan 2021