

## New England Biolabs Certificate of Analysis

Product Name: Monarch® RNA Purification Columns

Catalog Number: T2007L
Packaging Lot Number: 10247117
Expiration Date: 07/2026
Storage Temperature: 25°C

Specification Version: PS-T2007L v2.0

Monarch® RNA Purification Columns Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
T2118-1	Monarch® Spin Collection Tubes	10234057	Pass	
T2007-1	Monarch® RNA Purification Columns	10246539	Pass	

Assay Name/Specification	Lot # 10247117
Functional Testing (DNA Recovery and Purity)	Pass
Twenty-four Monarch® RNA Purification Columns are tested with 10 µg of Lambda/BstEII	
Marker resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are	
≥1.75 in ≥80% of the samples.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Julie Gaumont
Production Scientist
01 Aug 2024

Julie Gaumont

Packaging Quality Control Inspector

08 Aug 2024



T2007L / Lot: 10247117

Page 1 of 1