


New England Biolabs Certificate of Analysis

Product Name: *Monarch gDNA Removal Columns*
 Catalog Number: *T2017L*
 Packaging Lot Number: *10053945*
 Expiration Date: *10/2021*
 Storage Temperature: *25°C*
 Specification Version: *PS-T2017L v1.0*

Monarch gDNA Removal Columns Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T2018-1	Monarch® Collection Tubes II (50)	10054029	Pass
T2017-1	Monarch gDNA Removal Columns	10054005	Pass

Assay Name/Specification	Lot # 10053945
Functional Testing (DNA Recovery) Eight Monarch® gDNA Removal Columns are tested with 10 µg of Lambda/BstEII Marker resulting in ≥65% recovery in ≥80% of the samples.	Pass

This product has been tested and shown to be in compliance with all specifications.



Ana Egana
Production Scientist
11 May 2020



Heidi Church
Packaging Quality Control Inspector
11 May 2020