

## New England Biolabs Certificate of Analysis

Product Name: *Monarch® gDNA Removal Columns*  
 Catalog Number: *T2017L*  
 Packaging Lot Number: *10247134*  
 Expiration Date: *07/2026*  
 Storage Temperature: *25°C*  
 Specification Version: *PS-T2017L v2.0*

Monarch® gDNA Removal Columns Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T2118-1	Monarch® Spin Collection Tubes	10234057	Pass
T2017-1	Monarch® gDNA Removal Columns	10250040	Pass

Assay Name/Specification	Lot # 10247134
<b>Functional Testing (DNA Recovery)</b> Eight Monarch® gDNA Removal Columns are tested with 10 µg of Lambda/BstEII Marker resulting in ≥65% recovery in ≥80% of the samples.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



Julie Gaumont  
Production Scientist  
01 Aug 2024



Julie Gaumont  
Packaging Quality Control Inspector  
02 Aug 2024