

New England Biolabs Certificate of Analysis

Product Name: *Monarch® Spin Columns S1A and Tubes*
 Catalog Number: *T2037L*
 Packaging Lot Number: *10237541*
 Expiration Date: *01/2026*
 Storage Temperature: *25°C*
 Specification Version: *PS-T2037L v3.0*

Monarch® Spin Columns S1A and Tubes Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T2037-1	Monarch® Spin Columns S1A	10207583	Pass
T2018-1	Monarch® Collection Tubes II	10176509	Pass


Assay Name/Specification	Lot # 10237541
Functional Testing (RNA Recovery and Purity) Twenty-four Monarch® RNA Cleanup Columns (T2037, 10 µg) are tested with 1 µg of RNA resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Darcie Spaulding
Production Scientist
11 Jun 2024



Julie Gaumont
Packaging Quality Control Inspector
11 Jun 2024