

## New England Biolabs Certificate of Analysis

**Product Name:** Monarch® Spin Columns S1A and Tubes  
**Catalog Number:** T2037L  
**Packaging Lot Number:** 10287227  
**Expiration Date:** 02/2027  
**Storage Temperature:** 25°C  
**Specification Version:** PS-T2037L v4.0

Monarch® Spin Columns S1A and Tubes Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T2118-1	Monarch® Spin Collection Tubes	10245009	Pass
T2037-1	Monarch® Spin Columns S1A	10276074	Pass

Assay Name/Specification	Lot # 10287227
<b>Functional Testing (DNA Recovery and Purity)</b> Twenty-four Monarch® Spin Columns S1A are tested with 1 µg of 1 kb DNA Ladder resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	<b>Pass</b>
<b>Functional Testing (RNA Recovery and Purity)</b> Twenty-four Monarch® Spin Columns S1A are tested with 1 µg of RNA resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



Darcie Spaulding  
Production Scientist  
14 Apr 2025



Julie Gaumont  
Packaging Quality Control Inspector  
14 Apr 2025