


## New England Biolabs Certificate of Analysis

Product Name: Monarch™ RNA (50ug) Columns & Tubes  
 Catalog Number: T2047L  
 Lot Number: 0011806  
 Expiration Date: 00/0000  
 Storage Temperature: 25°C  
 Specification Version: PS-T2047L v1.0

Assay Name/Specification	Lot # 0011806
<b>Functional Testing (RNA Recovery and Purity)</b> Twenty-four Monarch® RNA Cleanup Columns (T2047, 50 µg) are tested with 5 µg of rRNA resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.



Ana Egana  
 Production Scientist  
 20 Jul 2018



Heidi Church  
 Packaging Quality Control Inspector  
 20 Jul 2018