320 Newburyport Turnpike Rowley, MA 01969-2002 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Monarch® gDNA Elution Buffer

Catalog Number: T3016L
Packaging Lot Number: 10190082
Expiration Date: 06/2025
Storage Temperature: 25°C

Specification Version: PS-T3016S/L v2.0

Monarch® gDNA Elution Buffer Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T3016-2	Monarch® gDNA Elution Buffer	10195504	Pass

Assay Name/Specification	Lot # 10190082
Functional Testing (DNA Recovery and Purity)	Pass
Twenty-four Monarch® gDNA Purification Columns are tested with 5 µg of HindIII	
digested Lambda resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and	
260/230 are ≥1.75 in ≥80% of the samples.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Nikki \$amuelson Production Scientist

28 Jun 2023

Jviie Gaumont

Packaging Quality Control Inspector

09 Aug 2023



T3016L / Lot: 10190082

Page 1 of 1