

## New England Biolabs Certificate of Analysis


Product Name: *Monarch® HMW DNA Extraction Kit for Cells & Blood*  
 Catalog Number: *T3050S*  
 Packaging Lot Number: *10193013*  
 Expiration Date: *05/2025*  
 Storage Temperature: *25°C*  
 Specification Version: *PS-T3050S/L v2.0*

Monarch® HMW DNA Extraction Kit for Cells & Blood Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T3056-21	Monarch® gDNA Elution Buffer II	10195508	Pass
T3055-21	Monarch® Precipitation Enhancer	10186137	Pass
T3053-21	Monarch® gDNA Nuclei Lysis Buffer	10186133	Pass
T3052-21	Monarch® gDNA Nuclei Prep Buffer	10186125	Pass
T3051-21	Monarch® RBC Lysis Buffer	10186069	Pass
T3018-21	Monarch® RNase A	10184644	Pass
T3015-21	Monarch® gDNA Wash Buffer	10186061	Pass
T3005-21	Monarch® DNA Capture Beads (10)	10186078	Pass
T3004-21	Monarch® Bead Retainers (5)	10142461	Pass
T3003-21	Monarch® 2 ml Tubes (10)	10179463	Pass
T2018-23	Monarch® Collection Tubes II	10142459	Pass
P8107GVIAL	Proteinase K, Molecular Biology Grade	10198925	Pass

Assay Name/Specification	Lot # 10193013
<p><b>* Individual Product Component Note</b>            Standard Quality Control Tests are performed for each component included in Monarch® HMW DNA Extraction Kit for Cells &amp; Blood and meet the designated specifications.</p>	Pass
<p><b>Functional Testing (Genomic DNA Yield and Integrity Analysis) (Blood)</b>            Genomic DNA is purified from 8 samples of whole pig blood using the High Molecular Weight DNA Extraction from Blood (Frozen Mammalian Blood) protocol. Each set of reagents is functionally tested to ensure successful isolation of gDNA from frozen pig blood samples. Yield of genomic DNA is compared across various metrics including a minimum gDNA yield of 6 µg per extraction and a recovery equivalence of ≥85% as compared to the control lot. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples, and DIN values are ≥9 in ≥80% of the samples. In ≥ 80% of samples, ≥ 75% of the DNA is 50 kb in size or greater.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



---

Ana Egana  
Production Scientist  
09 Aug 2023



---

Julie Gaumont  
Packaging Quality Control Inspector  
09 Aug 2023