

New England Biolabs Certificate of Analysis

Product Name: NEBuffer™ 3
Catalog #: B7003S

Concentration: 10X Concentrate

 Lot #:
 0101802

 Assay Date:
 02/2018

 Expiration Date:
 02/2021

 Storage Temp:
 -20°C

Composition (1X): 100 mM NaCl, 50 mM Tris-HCl, 10 mM MgCl₂, 1 mM DTT, (pH 7.9 @ 25°C)

Specification Version: PS-B7003S v1.0 Effective Date: 30 Jan 2018

Assay Name/Specification (minimum release criteria)	Lot #0101802
Conductivity (buffers/solutions) - The conductivity of 10X NEBuffer 3 is between 85 and 115 mS at 25°C.	Pass
Endonuclease Activity (Nicking, Buffer) - A 50 μl reaction in 1X NEBuffer 3 containing 1 μg of supercoiled PhiX174 DNA incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Functional Testing (Restriction Digest, BSA, Buffer) - A 50 μl reaction in 1X NEBuffer 3 plus 100 μg/ml Bovine Serum Albumin containing 1 μg of pBC4 DNA and 1 unit of NotI incubated for 1 hour at 37°C results in complete digestion of the substrate DNA as determined by agarose gel electrophoresis.	Pass
Functional Testing (Restriction Digest, BSA, Buffer) - A 50 μl reaction in 1X NEBuffer 3 plus 100 μg/ml Bovine Serum Albumin containing 1 μg of Lambda DNA and 1 unit of AseI incubated for 1 hour at 37°C results in complete digestion of the substrate DNA as determined by agarose gel electrophoresis.	Pass
Non-Specific DNase Activity (16 hour, Buffer) - A 50 µl reaction in 1X NEBuffer 3 containing 1 µg of PhiX174-HaeIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
pH (buffers/solutions) - The pH of 10X NEBuffer 3 is between pH 7.8 and 8.0 at 25°C.	Pass
RNase Activity (Buffer) - A 10 µl reaction in 1X NEBuffer 3 containing 40 ng of a 300 base single-stranded RNA is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by fluorescent detection.	Pass

Authorized by Derek Robinson 30 Jan 2018







Inspected by
Tony Spear-Alfonso

20 Feb 2018