

New England Biolabs Certificate of Analysis

Product Name: *HiScribe™ T7 High Yield RNA Synthesis Kit*
 Catalog Number: *E2040S*
 Lot Number: *10014774*
 Expiration Date: *01/2020*
 Storage Temperature: *-20°C*
 Specification Version: *PS-E2040S v1.0*

HiScribe™ T7 High Yield RNA Synthesis Kit Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
N0454AVIAL	CTP Solution	0111802	Pass
N0453AVIAL	UTP Solution	0111802	Pass
N0452AVIAL	GTP Solution	0121805	Pass
N0451AVIAL	ATP Solution	0111802	Pass
N0426AVIAL	Fluc Control Template	0061803	Pass
M0255AVIAL	T7 RNA Polymerase Mix	10013740	Pass
B2041AVIAL	10X T7 Reaction Buffer	0151803	Pass

Assay Name/Specification	Lot # 10014774
Functional Testing (RNA Synthesis, Yield) A 20 µl reaction in 1X T7 Reaction Buffer in the presence of 40 mM NTPs containing 1 µg FLuc Control Template with 2 µl of the T7 RNA Polymerase Mix incubated for 2 hours at 37°C results in a minimum of 150 µg of a 1.8 kb RNA as determined by OD 260 and agarose gel electrophoresis.	Pass
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in HiScribe™ T7 High Yield RNA Synthesis Kit and meet the designated specifications.	Pass

This product has been tested and shown to be in compliance with all specifications.

Dongxian Yue

Dongxian Yue
Production Scientist
10 Jul 2018

Michael Tonello

Michael Tonello
Packaging Quality Control Inspector
19 Jul 2018