

## New England Biolabs Certificate of Analysis

Product Name: NEBNext<sup>®</sup> Globin & rRNA Depletion Kit (Human/Mouse/Rat)  
 Catalog Number: E7750L  
 Packaging Lot Number: 10068080  
 Expiration Date: 05/2021  
 Storage Temperature: -20°C  
 Specification Version: PS-E7750S/L/X v1.0

NEBNext <sup>®</sup> Globin & rRNA Depletion Kit (Human/Mouse/Rat) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7753-3VIAL	NEBNext <sup>®</sup> DNase I	10042874	Pass
E7752-3VIAL	NEBNext <sup>®</sup> Thermostable RNase H	10042875	Pass
E7751-3VIAL	NEBNext <sup>®</sup> Globin & rRNA Depletion Solutio	10043092	Pass
E6317-3VIAL	Nuclease-free Water	10046662	Pass
E6315-3VIAL	DNase I Reaction Buffer	10042877	Pass
E6314-3VIAL	NEBNext <sup>®</sup> Probe Hybridization Buffer	10042879	Pass
E6312-3VIAL	RNase H Reaction Buffer	10042876	Pass

Assay Name/Specification	Lot # 10068080
<p><b>* Individual Product Component Note</b>            Standard Quality Control Tests are performed for each component included in NEBNext<sup>®</sup> Globin &amp; rRNA Depletion Kit (Human/Mouse/Rat) and meet the designated specifications.</p>	Pass
<p><b>Functional Testing (Globin and rRNA Depletion)</b>            The NEBNext<sup>®</sup> Globin &amp; rRNA Depletion Kit (Human/Mouse/Rat) is functionally validated using commercially available human RNA extracted from blood. After treatment with the kit, a high throughput sequencing library is made and sequenced. Reads from this library are classified as being derived from ribosomal RNA based on sequence homology. This method produces libraries with less than 10% rRNA and less than 5% globin RNA.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

*Christine Sumner*

---

Christine Sumner  
Production Scientist  
13 Feb 2020

*Michael Tonello*

---

Michael Tonello  
Packaging Quality Control Inspector  
13 Feb 2020