
New England Biolabs Certificate of Analysis

Product Name: 2-Log DNA Ladder (0.1-10.0 kb)
Catalog #: N3200S/L
Concentration: 1 mg/ml
Unit Definition: N/A
Lot #: 0791510
Assay Date: 10/2015
Expiration Date: 10/2017
Storage Temp: -20°C
Storage Conditions: 10 mM Tris-HCl (pH 8.0), 1 mM EDTA
Storage Buffer: 10 mM Tris-HCl (pH 8.0), 1 mM EDTA
Specification Version: PS-N3200S/L v1.0
Effective Date: 08 Oct 2015



New England Biolabs Certificate of Analysis

Assay Name/Specification (minimum release criteria)	Lot #0791510
A260/A280 Assay - The ratio of UV absorption of 2-Log DNA Ladder (0.1-10.0 kb) at 260 and 280 nm is between 1.8 and 2.0.	Pass
DNA Concentration (A260) - The concentration of 2-Log DNA Ladder (0.1-10.0 kb) is between 1000 and 1050 µg/ml as determined by UV absorption at 260 nm.	Pass
Electrophoretic Pattern (Marker) - The banding pattern of 2-Log DNA Ladder (0.1-10.0 kb) on a 1.2% agarose gel shows discrete, clearly identifiable bands at each band of the marker, when stained with Ethidium Bromide at a concentration of 0.5 µg/ml.	Pass
Non-Specific DNase Activity (DNA, 16 hour) - A 50 µl reaction in 1X NEBuffer 2 containing 5 µg of 2-Log DNA Ladder (0.1-10.0 kb) incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass



Authorized by
Derek Robinson
08 Oct 2015



Inspected by
Vanessa Mathieu-Sheltry
01 Dec 2015

