240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: ScrFI

Catalog #: R0110S/L
Concentration: 5,000 units/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction

volume of 50 μ l.

 Lot #:
 0111511

 Assay Date:
 11/2015

 Expiration Date:
 11/2017

 Storage Temp:
 -20°C

Storage Buffer: 250 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 0.15% Triton X-100, 200 μ g/ml

BSA

Specification Version: PS-R0110S/L v1.0
Effective Date: 29 May 2013

| Assay Name/Specification (minimum release criteria) | Lot #0111511 |
|---|--------------|
| Exonuclease Activity (Radioactivity Release) - A 50 μl reaction in CutSmart TM Buffer containing 1 μg of a mixture of single and double-stranded [³ H] <i>E. coli</i> DNA and a minimum of 15 units of ScrFI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity. | Pass |
| Ligation and Recutting (Terminal Integrity) - After a 2-fold over-digestion of Lambda DNA with ScrFI, <5% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with ScrFI. | Pass |
| Non-Specific DNase Activity (16 hour) - A 50 μ l reaction in CutSmart TM Buffer containing 1 μ g of Lambda DNA and a minimum of 5 Units of ScrFI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. NOTE: although no nuclease degradation is detected under these conditions, extended incubations and/or high concentrations of this enzyme may result in star activity. See the product FAQ for recommended reaction conditions for this enzyme. | Pass |
| Protein Purity Assay (SDS-PAGE) - ScrFI is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection. | Pass |

^{*} The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.

M.W. Southworth

Authorized by Maurice Southworth 29 May 2013







Inspected by Stephanie Cornelio 30 Dec 2015

Stephani Unetto