NEW ENGLAND BioLabs® Inc.

240 County Road
Ipswich, MA 01938-2723

Tel 978-927-5054
Fax 978-921-1350

www.neb.com
info@neb.com

New England Biolabs
Certificate of Analysis

Product Name:	*SnaBI*
Catalog #:	*R0130M*
Concentration:	*25,000 units/ml*
Unit Definition:	*One unit is defined as the amount of enzyme required to digest 1 µg of T7 DNA in 1 hour at 37°C in a total reaction volume of 50 µl.*
Lot #:	*0471607*
Assay Date:	*07/2016*
Expiration Date:	*07/2018*
Storage Temp:	*-20°C*
Storage Conditions:	*50 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA*
Specification Version:	*PS-R0130M v1.0*
Effective Date:	*05 Sep 2013*

Assay Name/Specification (minimum release criteria)	**Lot #0471607**
Endonuclease Activity (Nicking) - A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 5 Units of SnaBI incubated for 4 hours at 37°C results in <20% conversion to the nicked form as determined by agarose gel electrophoresis.	**Pass**
Exonuclease Activity (Radioactivity Release) - A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [^{3}H] *E. coli* DNA and a minimum of 50 units of SnaBI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	**Pass**
Ligation and Recutting (Terminal Integrity) - After a 20-fold over-digestion of T7 DNA DNA with SnaBI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with SnaBI.	**Pass**
Non-Specific DNase Activity (16 Hour) - A 50 µl reaction in CutSmart™ Buffer containing 1 µg of T7 DNA and a minimum of 5 units of SnaBI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	**Pass**

* *The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.*

Authorized by
Derek Robinson
05 Sep 2013

Inspected by
Stephanie Cornelio
06 Jul 2016

ISO 9001 Registered Quality Management
ISO 14001 Registered Environmental Management
ISO 13485 Registered Medical Devices